

**RIH – UPPER EXTREMITY RUNOFF CTA
SIEMENS DEFINITION AS20 PROTOCOL**

Position/Landmark	Head first or feet first-Supine. The arm should be placed over the patient's head when possible. Zero appropriately																
Topogram Direction	Craniocaudal																
Respiratory Phase	Suspension																
Scan Type	Helical																
Ref kV/Ref mAs/Rotation time (sec) Pitch / Speed (mm/rotation) Safire Strength / Dose Optimization	Care kV 120 / Care Dose4D 180 /0.5sec 1.2:1 , 15.00mm 3 / 11																
Detector width x Rows = Beam Collimation	0.625mm x 20 = 12.5mm																
Average Tube Output	ctdi – 8.1 mGy dlp – 330 mGy.cm																
Helical Set Slice Thickness/ Spacing Algorithm Recon Destination	<table border="0"> <tr> <td></td> <td>body</td> <td>thickness/</td> <td>recon</td> </tr> <tr> <td></td> <td>part</td> <td>spacing</td> <td>destination</td> </tr> <tr> <td>1</td> <td>run-off ct angio</td> <td>2mm x 2mm</td> <td>I31f med smooth pacs</td> </tr> <tr> <td>2</td> <td>thin ct angio</td> <td>.75mm x .7mm</td> <td>I31f med smooth mpr/TereRecon</td> </tr> </table>		body	thickness/	recon		part	spacing	destination	1	run-off ct angio	2mm x 2mm	I31f med smooth pacs	2	thin ct angio	.75mm x .7mm	I31f med smooth mpr/TereRecon
	body	thickness/	recon														
	part	spacing	destination														
1	run-off ct angio	2mm x 2mm	I31f med smooth pacs														
2	thin ct angio	.75mm x .7mm	I31f med smooth mpr/TereRecon														
Scan Start / End Locations	determined by technologist or radiologist to include the anatomy of interest																
DFOV	18cm decrease appropriately																
IV Contrast Volume / Type / Rate	100mL Iohexol (Omnipaque 350) / 4mL per second																
Scan Delay	Smart Prep at aortic arch or proximal extremity																
2D/3D Technique Used	3mm x 3mm sagittal and coronal upper extremity , mip mode manually transferred to PACS. 3d run-off ct angiogram, manually transferred to PACS. Thick run-off mip rotation , manually transferred to PACS.																
Comments: The cta is done using bolus tracking at the level of the aortic arch. The threshold trigger is +100 HU. Recon 2 is thin for reformats. 3mm x 3mm coronal reformats, mip mode region are created from this helical image data set. Thick mip rotation of the arterial anatomy.																	
Images required in PACS	Topograms, 2mm x 2mm axial run-off cta, 3mm x 3mm coronal upper extremity cta, 3mm x 3mm sagittal upper extremity cta, 3d run-off ct angiogram rotation, Thick mip rotation of the arterial anatomy. Patient Protocol																